



Temperature Log for Refrigerator – Fahrenheit

DAYS 1-15

Month/Year _____ VFC or VFAAR PIN _____

Facility Name _____ Room _____

Monitor temperatures closely!

1. Monitor storage temperatures with a certified, calibrated digital data logger.
2. Write your initials below in "Staff Initials," and note the time in "Exact Time."
3. Record temps twice each workday.
4. Record the min/max temps once each workday – preferably in the morning.
5. Put an "X" in the row that corresponds to the refrigerator's temperature.
6. If any out-of-range temp, see instructions to the right.
7. After each month has ended, save each month's log for 3 years.

Take action if temp is out of range – too warm (above 46°F) or too cold (below 35°F).

1. Label exposed vaccine "do not use," and store it under proper conditions as quickly as possible. Do not discard vaccines unless directed by the Philadelphia Immunization Program.
2. Record the out-of-range temps and the room temp in the "Action" area on the bottom of the log.
3. Notify your vaccine coordinator and contact TempCheck (TempCheck@phila.gov, 215-685-6777).
4. Document all action taken.

Day of Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
Staff Initials	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Exact Time	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm
	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Min/Max Temp (since previous reading)	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••
Danger! Temperatures above 46°F are too warm! Write any out-of-range temps and room temp on the lines below and call the Immunization Program immediately!															
temperatures	46°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	45°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	44°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	43°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	42°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	41°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	Ideal temp → 40°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
in-range	39°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	38°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	37°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	36°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Danger! Temperatures below 36° F are too cold. Write any out-of-range temps and room temp on the line below and call the Immunization Program immediately!															
action	Write any out-of-range temps (above 46°F or below 36°F) here.	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	Room Temperature	:	:	:	:	:	:	:	:	:	:	:	:	:	:

If you have a vaccine storage issue, contact us immediately:

TempCheck: 215-685-6777

VFC/VFAAR Coordinator: 215-685-6424

2018

<https://kids.phila.gov>

FAX for vaccine orders: 215-238-6939

FAX for storage issues: 215-238-6948



Temperature Log for Refrigerator – Fahrenheit

DAYS 16–31

Month/Year _____ VFC or VFAAR PIN _____

Facility Name _____ Room _____

Monitor temperatures closely!

1. Monitor storage temperatures with a certified, calibrated digital data logger.
2. Write your initials below in "Staff Initials," and note the time in "Exact Time."
3. Record temps twice each workday.
4. Record the min/max temps once each workday – preferably in the morning.
5. Put an "X" in the row that corresponds to the refrigerator's temperature.
6. If any out-of-range temp, see instructions to the right.
7. After each month has ended, save each month's log for 3 years.

Take action if temp is out of range – too warm (above 46°F) or too cold (below 35°F).

1. Label exposed vaccine "do not use," and store it under proper conditions as quickly as possible. Do not discard vaccines unless directed by the Philadelphia Immunization Program.
2. Record the out-of-range temps and the room temp in the "Action" area on the bottom of the log.
3. Notify your vaccine coordinator and contact TempCheck (TempCheck@phila.gov, 215-685-6777).
4. Document all action taken.

Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Staff Initials	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Exact Time	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm
	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Min/Max Temp (since previous reading)	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••	•••••
Danger! Temperatures above 5°F are too warm! Write any out-of-range temps and room temp on the lines below and call the Immunization Program immediately!																
temperatures	46°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	45°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	44°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	43°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	42°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	41°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	40°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
in-range	39°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	38°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	37°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	36°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	Danger! Temperatures below 36 F are too cold. Write any out-of-range temps and room temp on the line below and call the Immunization Program immediately!															
action	Write any out-of-range temps (above 46°F or below 36°F) here.	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	Room Temperature	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:

If you have a vaccine storage issue, contact us immediately:

TempCheck: 215-685-6777

VFC/VFAAR Coordinator: 215-685-6424

2018

<https://kids.phila.gov>

FAX for vaccine orders: 215-238-6939

FAX for storage issues: 215-238-6948